



Certificate of Calibration

Metrology Calibration laboratory

Certificate no.: E 45264

Device under test

Description:	Compact Temperature Calibrator		
Serial number:	656874-00299	Model:	CTC-155 A
			
Manufacturer:	AMETEK Denmark A/S		
Range:	-39°C to 155°C-38,2°F to 311°F		
Laboratory subject no:	SA 72395		
Date of calibration:	sep 22. 2017		

Customer information

Client:	AMETEK Denmark A/S	Number:	48168000
Address:	Gydevang 32-34 3450 Allerød	Phone:	+45 4816 8000

Remarks

The calibration was carried out with the laboratory reference sensor immersed 95 mm into the calibration insert for each temperature point. The insert hole where the reference probe was immersed, is equal to the probe diameter + 0,2 mm.
Standard insulation plug was mounted and present during the calibration
Contribution to reported combined uncertainties arises from unit stability and reference equipment including correction from immersion depth.

Calibrated by:

Ole Asklund Jørgensen
Calibration technician

Approved by:

Ole A. Jørgensen

Digitalt signeret af
Ole A. Jørgensen
Dato: 2017.09.28
10:13:30 +02'00'

A summary of this report may be issued only when it is clearly stated that it is a summary and only if the full report is cacheable to the public, or if the summary has been approved by AMETEK Denmark A/S, Metrology Laboratory.



Certificate of Calibration

Metrology Calibration laboratory

Certificate no.: E 45264

Calibration conditions

Calibration procedure:	126553	Reference id.:	Y007
Ambient temperature:	23 ±2 °C	Reference id.:	Y007
Relative humidity:	30..85%		

Reference equipment applied

Description:	Manufacturer:	Model:	Id. number:
Reference sensor	AMETEK Denmark A/S	162 CE	T142
Resistance Bridge	AMETEK Denmark A/S	DTI 1000 B	E058
Insert	AMETEK Denmark A/S	129332	IN180

Set value		Read value		True value		Deviation		Uncertainty	
°C	°F	°C	°F	°C	°F	°C	°F	°C	°F
-25,00	-13,00	-25,00	-13,00	-25,012	-13,022	0,012	0,022	±0,05	±0,09
-15,00	5,00	-15,00	5,00	-14,991	5,016	-0,009	-0,016	±0,05	±0,09
0,00	32,00	0,00	32,00	0,013	32,023	-0,013	-0,023	±0,05	±0,09
50,00	122,00	50,00	122,00	50,013	122,023	-0,013	-0,023	±0,05	±0,09
100,00	212,00	100,00	212,00	100,015	212,027	-0,015	-0,027	±0,05	±0,09
155,00	311,00	155,00	311,00	155,008	311,014	-0,008	-0,014	±0,05	±0,09

The stated uncertainty is based on the entire set-up including object under test.

The reported uncertainty is based upon a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximate 95%. The uncertainty evaluation has been carried out in accordance with DANAK requirements. The uncertainty is calculated following EA-4/02

DANAK

DANAK is the national accreditation body in Denmark in compliance with EU regulation No. 765/2008

DANAK participates in the multilateral agreements for testing and calibration under European co-operation for Accreditation (EA) and under International Laboratories Accreditation Cooperation (ILAC) based on peer evaluation. Accredited test reports and calibration certificates issued by laboratories accredited by DANAK are recognized cross border by members of EA and ILAC equal to test reports and calibration certificates issued by these members' accredited laboratories.

The use of the accreditation mark on test reports and calibration certificates or reference to accreditation, documents that the service is provided as an accredited service under the company's DANAK accreditation.

The calibration certificate is covered by DANAK accreditation and the multilateral agreements from EA and ILAC for calibration which ensures that measurements are traceable to the international system of units, SI.